

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE010423AL\
 Data File : FE039629.D
 Signal(s) : FID1B.ch
 Acq On : 04 Jan 2023 15:30
 Operator : YP\AJ
 Sample : PB150033BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 PB150033BL

Integration File: autoint1.e
 Quant Time: Jan 05 08:33:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 121722.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 17 15:09:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.926	6421229	43.028 ug/ml
Spiked Amount 50.000		Recovery =	86.06%
12) S 1-chlorooctadecane (S...	13.362	5064007	43.997 ug/ml
Spiked Amount 50.000		Recovery =	87.99%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE010423AL\
Data File : FE039629.D
Signal(s) : FID1B.ch
Acq On : 04 Jan 2023 15:30
Operator : YP\AJ
Sample : PB150033BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
PB150033BL

Integration File: autoint1.e
Quant Time: Jan 05 08:33:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 121722.M
Quant Title : GC Extractables
QLast Update : Sat Dec 17 15:09:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.926 min
Delta R.T.: -0.024 min
Response: 6421229
Conc: 43.03 ug/ml

Instrument :
FID_E
ClientSampleId :
PB150033BL



#12 1-chlorooctadecane (SURR)

R.T.: 13.362 min
Delta R.T.: -0.021 min
Response: 5064007
Conc: 44.00 ug/ml