

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE010722AL\
 Data File : FE034677.D
 Signal(s) : FID1B.CH
 Acq On : 07 Jan 2022 23:51
 Operator : AJ\MA
 Sample : N1067-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jan 08 02:37:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 010622.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 07 12:52:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.093	4055779	29.317 ug/ml
Spiked Amount 50.000		Recovery =	58.63%
12) S 1-chlorooctadecane (S...	12.526	3578222	33.984 ug/ml
Spiked Amount 50.000		Recovery =	67.97%
Target Compounds			
3) T A~Naphthalene (C11.7)	5.608f	123753	<MDL ug/ml
5) T A~2-methylnaphthalene...	6.641	111167	<MDL ug/ml
7) T n-Hexadecane (C16)	9.362	76972	<MDL ug/ml
10) T n-Eicosane (C20)	12.121	1283029	10.403 ug/ml
14) T n-Tetracosane (C24)	14.323f	255792	2.079 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE010722AL\
Data File : FE034677.D
Signal(s) : FID1B.CH
Acq On : 07 Jan 2022 23:51
Operator : AJ\MA
Sample : N1067-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jan 08 02:37:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 010622.M
Quant Title : GC Extractables
QLast Update : Fri Jan 07 12:52:40 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.093 min
Delta R.T.: -0.018 min
Response: 4055779
Conc: 29.32 ug/ml

Instrument :
FID_E
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.121 min
Delta R.T.: 0.020 min
Response: 1283029
Conc: 10.40 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.526 min
Delta R.T.: -0.017 min
Response: 3578222
Conc: 33.98 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.323 min
Delta R.T.: -0.070 min
Response: 255792
Conc: 2.08 ug/ml