

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE010919\
 Data File : FE029397.D
 Signal(s) : FID1B.CH
 Acq On : 09 Jan 2019 17:46
 Operator : AJ\SJ
 Sample : K1082-01MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 10 02:57:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE010419.M
 Quant Title :
 QLast Update : Fri Jan 04 15:40:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
13) S	TETRACOSANE-d50 (SURR...	14.832	3259158	16.996 ug/ml
Target Compounds				
1)	N-OCTANE	2.505	2024587	11.554 ug/ml
3)	N-DECANE	4.549	3019234	10.954 ug/ml
5)	N-DODECANE	6.562	3184531	17.024 ug/ml
6)	N-TETRADECANE	8.441	2008830	10.597 ug/ml
7)	N-HEXADECANE	9.967	3461320	18.145 ug/ml
8)	N-OCTADECANE	11.414	3512280	18.350 ug/ml
9)	N-EICOSANE	12.727	3530378	18.611 ug/ml
10)	N-TETRACOSANE	15.037	3565353	18.787 ug/ml
11)	N-OCTACOSANE	17.014	3620583	19.388 ug/ml
12)	N-DOTRIACONTANE	18.739	3373781	20.301 ug/ml
14)	N-HEXATRIACONTANE	20.264	2286410	22.442 ug/ml
15)	N-TETRACONTANE	21.990	1900723	21.899 ug/ml
16)	N-TETRTETRACOCTANE	24.906	102615	1.371 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE010919\
Data File : FE029397.D
Signal(s) : FID1B.CH
Acq On : 09 Jan 2019 17:46
Operator : AJ\SJ
Sample : K1082-01MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 10 02:57:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE010419.M
Quant Title :
QLast Update : Fri Jan 04 15:40:11 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

