

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE011320\
 Data File : FE031089.D
 Signal(s) : FID1B.CH
 Acq On : 13 Jan 2020 11:20
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Jan 14 07:30:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE010420.M
 Quant Title :
 QLast Update : Mon Jan 06 08:13:00 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.625	17142915	51.374 ug/ml
Target Compounds			
1) N-OCTANE	2.334	19283294	50.184 ug/ml
2) N-DECANE	4.368	20465250	54.467 ug/ml
3) N-DODECANE	6.376	20837335	58.246 ug/ml
4) N-TETRADECANE	8.171	20630195	53.953 ug/ml
5) N-HEXADECANE	9.773	20661374	53.369 ug/ml
6) N-OCTADECANE	11.216	20354120	52.906 ug/ml
7) N-EICOSANE	12.527	20284484	52.136 ug/ml
8) N-DOCOSANE	13.727	19956735	51.676 ug/ml
10) N-TETRACOSANE	14.832	19753487	51.502 ug/ml
11) N-HEXACOSANE	15.853	19499482	50.980 ug/ml
12) N-OCTACOSANE	16.804	19486342	50.689 ug/ml
13) N-TRIACONTANE	17.692	19345284	50.462 ug/ml
14) N-DOTRIACONTANE	18.524	18323704	50.663 ug/ml
15) N-TETRATRIACONTANE	19.308	17112014	51.318 ug/ml
16) N-HEXATRIACONTANE	20.046	15303702	51.728 ug/ml
17) N-OCTATRIACONTANE	20.774	13612630	52.498 ug/ml
18) N-TETRACONTANE	21.662	13123988	52.881 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE011320\
Data File : FE031089.D
Signal(s) : FID1B.CH
Acq On : 13 Jan 2020 11:20
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Jan 14 07:30:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE010420.M
Quant Title :
QLast Update : Mon Jan 06 08:13:00 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

