

Data Path : P:\HPCHEM1\FID E\Data\FE011518\
 Data File : FE025611.D
 Signal(s) : FID1B.CH
 Acq On : 15 Jan 2018 21:39
 Operator : UA/AJ
 Sample : J1019-09 10X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jan 16 00:22:36 2018
 Quant Method : P:\HPCHEM1\FID_E\methods\FE011218.M
 Quant Title :
 QLast Update : Fri Jan 12 18:22:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.831	261452	1.437 ug/ml
Target Compounds			
3) N-DODECANE	6.592	261334	1.638 ug/ml
4) N-TETRADECANE	8.385	437736	2.402 ug/ml
5) N-HEXADECANE	9.986	486940	2.384 ug/ml
6) N-OCTADECANE	11.426	384614	1.740 ug/ml
7) N-EICOSANE	12.736	226163	1.020 ug/ml
8) N-DOCOSANE	13.985	187158	0.853 ug/ml
10) N-TETRACOSANE	15.036	161003	0.755 ug/ml
14) N-DOTRIACONTANE	18.721	301996	1.862 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : P:\HPCHEM1\FID E\Data\FE011518\
Data File : FE025611.D
Signal(s) : FID1B.CH
Acq On : 15 Jan 2018 21:39
Operator : UA/AJ
Sample : J1019-09 10X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jan 16 00:22:36 2018
Quant Method : P:\HPCHEM1\FID_E\methods\FE011218.M
Quant Title :
QLast Update : Fri Jan 12 18:22:53 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE


R.T.: 6.592 min
Delta R.T.: -0.023 min
Response: 261334
Conc: 1.64 ug/ml

Instrument :
FID_E
ClientSampleId :




#4 N-TETRADECANE

R.T.: 8.385 min
Delta R.T.: -0.027 min
Response: 437736
Conc: 2.40 ug/ml




#5 N-HEXADECANE

R.T.: 9.986 min
Delta R.T.: -0.029 min
Response: 486940
Conc: 2.38 ug/ml



#6 N-OCTADECANE


R.T.: 11.426 min
Delta R.T.: -0.030 min
Response: 384614
Conc: 1.74 ug/ml



#7 N-EICOSANE


R.T.: 12.736 min
Delta R.T.: -0.031 min
Response: 226163
Conc: 1.02 ug/ml

Instrument :
FID_E
ClientSampleId :




#8 N-DOCOSANE

R.T.: 13.985 min
Delta R.T.: 0.021 min
Response: 187158
Conc: 0.85 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.831 min
Delta R.T.: -0.031 min
Response: 261452
Conc: 1.44 ug/ml



#10 N-TETRACOSANE

R.T.: 15.036 min
Delta R.T.: -0.030 min
Response: 161003
Conc: 0.75 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.721 min
Delta R.T.: -0.032 min
Response: 301996
Conc: 1.86 ug/ml

Instrument :
FID_E
ClientSampleId :