

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE011520\
 Data File : FE031132.D
 Signal(s) : FID1B.CH
 Acq On : 15 Jan 2020 17:33
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 16 00:28:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE010420.M
 Quant Title :
 QLast Update : Mon Jan 06 08:13:00 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.635	15711472	47.085 ug/ml
Target Compounds			
2) N-DECANE	4.371	19196031	51.089 ug/ml
3) N-DODECANE	6.381	19505574	55.267 ug/ml
4) N-TETRADECANE	8.176	19209845	50.239 ug/ml
5) N-HEXADECANE	9.779	19021278	49.133 ug/ml
6) N-OCTADECANE	11.223	18538597	48.187 ug/ml
7) N-EICOSANE	12.535	18652650	47.941 ug/ml
8) N-DOCOSANE	13.736	18268616	47.304 ug/ml
10) N-TETRACOSANE	14.841	18009608	46.955 ug/ml
11) N-HEXACOSANE	15.863	17783119	46.493 ug/ml
12) N-OCTACOSANE	16.815	17798718	46.299 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE011520\
Data File : FE031132.D
Signal(s) : FID1B.CH
Acq On : 15 Jan 2020 17:33
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 16 00:28:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE010420.M
Quant Title :
QLast Update : Mon Jan 06 08:13:00 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

