

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012423AL\
 Data File : FE040024.D
 Signal(s) : FID1B.ch
 Acq On : 25 Jan 2023 06:08
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 25 06:33:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 011023.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 10 04:24:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.901	3066231	19.967 ug/ml
Spiked Amount 50.000		Recovery =	39.93%
12) S 1-chlorooctadecane (S...	13.339	2348327	20.123 ug/ml
Spiked Amount 50.000		Recovery =	40.25%
Target Compounds			
1) T n-Nonane (C9)	3.301	2686673	21.166 ug/ml
2) T n-Decane (C10)	4.549	2640790	21.265 ug/ml
3) T A~Naphthalene (C11.7)	6.250	2900494	20.567 ug/ml
4) T n-Dodecane (C12)	6.711	2677559	21.297 ug/ml
5) T A~2-methylnaphthalene...	7.345	2709649	20.688 ug/ml
6) T n-Tetradecane (C14)	8.537	2717908	21.132 ug/ml
7) T n-Hexadecane (C16)	10.143	2796780	20.731 ug/ml
8) T n-Octadecane (C18)	11.586	2849423	20.335 ug/ml
10) T n-Eicosane (C20)	12.895	2784963	20.101 ug/ml
11) T n-Heneicosane (C21)	13.506	2787203	19.974 ug/ml
13) T n-Docosane (C22)	14.093	2769559	19.953 ug/ml
14) T n-Tetracosane (C24)	15.195	2782343	20.209 ug/ml
15) T n-Hexacosane (C26)	16.216	2686579	20.129 ug/ml
16) T n-Octacosane (C28)	17.166	2665030	20.472 ug/ml
17) T n-Tricontane (C30)	18.054	2698005	21.156 ug/ml
18) T n-Dotriacontane (C32)	18.887	2716509	21.488 ug/ml
19) T n-Tetratriacontane (C34)	19.671	2581924	20.895 ug/ml
20) T n-Hexatriacontane (C36)	20.411	2449362	20.775 ug/ml
21) T n-Octatriacontane (C38)	21.142	2349292	19.656 ug/ml
22) T n-Tetracontane (C40)	22.040	2435037	20.063 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012423AL\
Data File : FE040024.D
Signal(s) : FID1B.ch
Acq On : 25 Jan 2023 06:08
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 25 06:33:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 011023.M
Quant Title : GC Extractables
QLast Update : Tue Jan 10 04:24:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

