

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012425\
 Data File : FE052053.D
 Signal(s) : FID1B.ch
 Acq On : 24 Jan 2025 16:56
 Operator : YP\AJ
 Sample : Q1168-11
 Misc : 5 PPB WATER LOQ
 ALS Vial : 26 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 27 00:46:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012325.M
 Quant Title :
 QLast Update : Fri Jan 24 03:06:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.251	1725562	17.325 ug/ml
Target Compounds			
1) N-OCTANE	2.411	351741	4.125 ug/ml
2) N-DECANE	4.905	405867	4.435 ug/ml
3) N-DODECANE	7.032	433611	4.340 ug/ml
4) N-TETRADECANE	8.837	464355	4.564 ug/ml
5) N-HEXADECANE	10.430	480435	4.509 ug/ml
6) N-OCTADECANE	11.862	505768	4.511 ug/ml
7) N-EICOSANE	13.164	511707	4.594 ug/ml
8) N-DOCOSANE	14.356	499722	4.501 ug/ml
10) N-TETRACOSANE	15.454	498856	4.510 ug/ml
11) N-HEXACOSANE	16.471	498272	4.571 ug/ml
12) N-OCTACOSANE	17.417	492441	4.562 ug/ml
13) N-TRIACONTANE	18.302	506289	4.718 ug/ml
14) N-DOTRIACONTANE	19.132	498429	4.773 ug/ml
15) N-TETRATRIACONTANE	19.914	457547	4.872 ug/ml
16) N-HEXATRIACONTANE	20.652	404541	5.010 ug/ml
17) N-OCTATRIACONTANE	21.437	370011	4.959 ug/ml
18) N-TETRACONTANE	22.434	354492	4.814 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012425\
Data File : FE052053.D
Signal(s) : FID1B.ch
Acq On : 24 Jan 2025 16:56
Operator : YP\AJ
Sample : Q1168-11
Misc : 5 PPB WATER LOQ
ALS Vial : 26 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 27 00:46:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012325.M
Quant Title :
QLast Update : Fri Jan 24 03:06:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

