

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012425\
 Data File : FE052055.D
 Signal(s) : FID1B.ch
 Acq On : 24 Jan 2025 17:56
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 27 00:46:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012325.M
 Quant Title :
 QLast Update : Fri Jan 24 03:06:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.252	4700771	47.197 ug/ml
Target Compounds			
1) N-OCTANE	2.410	4040216	47.377 ug/ml
2) N-DECANE	4.905	4338981	47.418 ug/ml
3) N-DODECANE	7.033	4729110	47.331 ug/ml
4) N-TETRADECANE	8.839	4790207	47.079 ug/ml
5) N-HEXADECANE	10.432	5010975	47.025 ug/ml
6) N-OCTADECANE	11.865	5271352	47.015 ug/ml
7) N-EICOSANE	13.167	5233313	46.986 ug/ml
8) N-DOCOSANE	14.358	5214017	46.960 ug/ml
10) N-TETRACOSANE	15.456	5219001	47.183 ug/ml
11) N-HEXACOSANE	16.472	5147692	47.222 ug/ml
12) N-OCTACOSANE	17.419	5101591	47.264 ug/ml
13) N-TRIACONTANE	18.304	5066626	47.219 ug/ml
14) N-DOTRIACONTANE	19.134	4922607	47.144 ug/ml
15) N-TETRATRIACONTANE	19.915	4421905	47.089 ug/ml
16) N-HEXATRIACONTANE	20.652	3772102	46.712 ug/ml
17) N-OCTATRIACONTANE	21.437	3476192	46.592 ug/ml
18) N-TETRACONTANE	22.436	3401276	46.185 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012425\
Data File : FE052055.D
Signal(s) : FID1B.ch
Acq On : 24 Jan 2025 17:56
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 27 00:46:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012325.M
Quant Title :
QLast Update : Fri Jan 24 03:06:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

