

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012425\
 Data File : FE052057.D
 Signal(s) : FID1B.ch
 Acq On : 24 Jan 2025 18:56
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 27 00:47:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012325.M
 Quant Title :
 QLast Update : Fri Jan 24 03:06:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.252	4697203	47.161 ug/mlm
Target Compounds			
2) N-DECANE	4.905	4363927	47.691 ug/ml
3) N-DODECANE	7.034	4761854	47.659 ug/ml
4) N-TETRADECANE	8.839	4821846	47.390 ug/ml
5) N-HEXADECANE	10.432	5034180	47.243 ug/ml
6) N-OCTADECANE	11.865	5286285	47.148 ug/ml
7) N-EICOSANE	13.166	5236833	47.018 ug/ml
8) N-DOCOSANE	14.359	5214214	46.962 ug/ml
10) N-TETRACOSANE	15.456	5210407	47.106 ug/ml
11) N-HEXACOSANE	16.473	5154738	47.287 ug/ml
12) N-OCTACOSANE	17.420	5111499	47.355 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012425\
Data File : FE052057.D
Signal(s) : FID1B.ch
Acq On : 24 Jan 2025 18:56
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 27 00:47:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012325.M
Quant Title :
QLast Update : Fri Jan 24 03:06:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

