

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012519\
 Data File : FE029538.D
 Signal(s) : FID1B.CH
 Acq On : 26 Jan 2019 7:35
 Operator : AJ\SJ
 Sample : K1238-26
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 28 02:49:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE010419.M
 Quant Title :
 QLast Update : Fri Jan 04 15:40:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.818	4317718	22.516 ug/ml
Target Compounds			
6) N-TETRADECANE	8.432	1140803	6.018 ug/ml
10) N-TETRACOSANE	15.016	157473	0.830 ug/ml
11) N-OCTACOSANE	16.993	561022	3.004 ug/ml
12) N-DOTRIACONTANE	18.717	280751	1.689 ug/ml
14) N-HEXATRIACONTANE	20.245	68952	0.677 ug/ml
15) N-TETRACONTANE	21.961	66073	0.761 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012519\
Data File : FE029538.D
Signal(s) : FID1B.CH
Acq On : 26 Jan 2019 7:35
Operator : AJ\SJ
Sample : K1238-26
Misc :
ALS Vial : 28 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 28 02:49:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE010419.M
Quant Title :
QLast Update : Fri Jan 04 15:40:11 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um






#6 N-TETRADECANE

R.T.: 8.432 min
Delta R.T.: 0.008 min
Response: 1140803
Conc: 6.02 ug/ml




#10 N-TETRACOSANE

R.T.: 15.016 min
Delta R.T.: -0.041 min
Response: 157473
Conc: 0.83 ug/ml




#11 N-OCTACOSANE

R.T.: 16.993 min
Delta R.T.: -0.031 min
Response: 561022
Conc: 3.00 ug/ml




#12 N-DOTRIACONTANE

R.T.: 18.717 min
Delta R.T.: -0.029 min
Response: 280751
Conc: 1.69 ug/ml




#13 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.818 min
Delta R.T.: -0.022 min
Response: 4317718
Conc: 22.52 ug/ml



#14 N-HEXATRIACONTANE

R.T.: 20.245 min
Delta R.T.: -0.023 min
Response: 68952
Conc: 0.68 ug/ml



#15 N-TETRACONTANE

R.T.: 21.961 min
Delta R.T.: -0.036 min
Response: 66073
Conc: 0.76 ug/ml