

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012525\
 Data File : FE052087.D
 Signal(s) : FID1B.ch
 Acq On : 25 Jan 2025 11:00
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Jan 27 02:39:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012325.M
 Quant Title :
 QLast Update : Fri Jan 24 03:06:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.249	4988679	50.088 ug/ml
Target Compounds			
2) N-DECANE	4.903	4625009	50.544 ug/ml
3) N-DODECANE	7.031	5041700	50.460 ug/ml
4) N-TETRADECANE	8.836	5119117	50.312 ug/ml
5) N-HEXADECANE	10.429	5361536	50.315 ug/ml
6) N-OCTADECANE	11.862	5647111	50.366 ug/ml
7) N-EICOSANE	13.164	5586431	50.157 ug/ml
8) N-DOCOSANE	14.355	5552022	50.004 ug/ml
10) N-TETRACOSANE	15.453	5539161	50.078 ug/ml
11) N-HEXACOSANE	16.468	5450665	50.002 ug/ml
12) N-OCTACOSANE	17.414	5395007	49.982 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE012525\
Data File : FE052087.D
Signal(s) : FID1B.ch
Acq On : 25 Jan 2025 11:00
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Jan 27 02:39:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012325.M
Quant Title :
QLast Update : Fri Jan 24 03:06:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

