

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE013120\
 Data File : FE031267.D
 Signal(s) : FID1B.CH
 Acq On : 31 Jan 2020 17:36
 Operator : DD\AJ
 Sample : PB126493BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 PB126493BS

Manual Integrations
 APPROVED

Sohil
 2/3/2020 2:37:08 PM

Integration File: autoint1.e
 Quant Time: Jan 31 23:33:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012920.M
 Quant Title :
 QLast Update : Thu Jan 30 04:09:51 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.614	4036998	12.415 ug/ml
Target Compounds			
1) N-OCTANE	2.323	4367780	11.712 ug/ml
2) N-DECANE	4.348	4671813	13.085 ug/ml
3) N-DODECANE	6.356	4724970	14.325 ug/ml
4) N-TETRADECANE	8.152	4826152	13.330 ug/ml
5) N-HEXADECANE	9.755	4826764	13.180 ug/ml
6) N-OCTADECANE	11.200	4831920	13.176 ug/ml
7) N-EICOSANE	12.512	4875094	13.124 ug/ml
8) N-DOCOSANE	13.713	4840842	13.007 ug/ml
10) N-TETRACOSANE	14.819	4778898	12.763 ug/ml
11) N-HEXACOSANE	15.842	4696175	12.600 ug/ml
12) N-OCTACOSANE	16.794	4676636	12.481 ug/ml
13) N-TRIACONTANE	17.684	4568280	12.210 ug/ml
14) N-DOTRIACONTANE	18.518	4400622	12.436 ug/ml
15) N-TETRATRIACONTANE	19.301	4023881	13.166 ug/mlm
16) N-HEXATRIACONTANE	20.042	3519221	13.879 ug/ml
17) N-OCTATRIACONTANE	20.771	3197339	16.075 ug/ml
18) N-TETRACONTANE	21.662	3084893	16.723 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE013120\
Data File : FE031267.D
Signal(s) : FID1B.CH
Acq On : 31 Jan 2020 17:36
Operator : DD\AJ
Sample : PB126493BS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
PB126493BS

Manual Integrations
APPROVED

Sohil
2/3/2020 2:37:08 PM

Integration File: autoint1.e
Quant Time: Jan 31 23:33:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE012920.M
Quant Title :
QLast Update : Thu Jan 30 04:09:51 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

