

Method Path : P:\HPCHEM1\FID_D\methods\
 Method File : FD020518.M
 Title :
 Last Update : Mon Feb 05 16:59:17 2018
 Response Via : Initial Calibration

Calibration Files

50 =FD019854.D 100 =FD019853.D 20 =FD019855.D
 10 =FD019856.D 5 =FD019857.D

	Compound	50	100	20	10	5	Avg		%RSD
1)	N-OCTANE	7.588	7.983	7.773	8.059	7.761	7.833	E4	2.41
2)	N-DECANE	7.895	8.240	7.968	8.260	8.001	8.073	E4	2.06
3)	N-DODECANE	8.152	8.482	8.162	8.485	8.289	8.314	E4	1.97
4)	N-TETRADECANE	8.396	8.650	8.348	8.794	8.654	8.568	E4	2.21
5)	N-HEXADECANE	8.364	8.509	8.398	8.972	9.001	8.649	E4	3.62
6)	N-OCTADECANE	8.329	8.388	8.593	9.405	9.476	8.838	E4	6.33
7)	N-EICOSANE	8.058	8.094	8.276	8.972	9.400	8.560	E4	6.98
8)	N-DOCOSANE	7.858	7.908	7.921	8.477	8.545	8.142	E4	4.16
9) S	TETRACOSANE-d50 (S...	7.006	6.981	6.866	7.452	7.645	7.190	E4	4.71
10)	N-TETRACOSANE	7.921	7.940	7.786	8.433	8.322	8.080	E4	3.47
11)	N-HEXACOSANE	8.038	8.071	7.722	8.441	8.587	8.172	E4	4.22
12)	N-OCTACOSANE	8.178	8.246	7.812	8.794	9.138	8.434	E4	6.26
13)	N-TRIACONTANE	8.396	8.500	8.194	9.448	9.756	8.859	E4	7.86
14)	N-DOTRIACONTANE	8.105	8.168	8.136	9.363	9.559	8.666	E4	8.41
15)	N-TETRATRIACONTANE	7.306	7.233	7.234	8.280	8.798	7.770	E4	9.34
16)	N-HEXATRIACONTANE	5.836	5.894	5.852	6.549	6.623	6.151	E4	6.48
17)	N-OCTATRIACONTANE	5.545	5.571	5.620	6.457	6.380	5.915	E4	7.80
18)	N-TETRACONTANE	5.432	5.494	5.460	6.150	6.225	5.752	E4	6.93

(#) = Out of Range