

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE021022AL\  
 Data File : FE034966.D  
 Signal(s) : FID1B.CH  
 Acq On : 10 Feb 2022 15:55  
 Operator : YP\AJ  
 Sample : PB142582BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :

Integration File: autoint1.e  
 Quant Time: Feb 11 00:40:36 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 020822.M  
 Quant Title : GC Extractables  
 QLast Update : Tue Feb 08 22:05:03 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.128 | 12611091   | 43.060 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 86.12%       |
| 12) S 1-chlorooctadecane (S... | 12.557 | 9795711    | 43.813 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 87.63%       |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE021022AL\  
Data File : FE034966.D  
Signal(s) : FID1B.CH  
Acq On : 10 Feb 2022 15:55  
Operator : YP\AJ  
Sample : PB142582BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :

Integration File: autoint1.e  
Quant Time: Feb 11 00:40:36 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 020822.M  
Quant Title : GC Extractables  
QLast Update : Tue Feb 08 22:05:03 2022  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.128 min  
Delta R.T.: -0.004 min  
Response: 12611091  
Conc: 43.06 ug/ml

Instrument :  
FID\_E  
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 12.557 min  
Delta R.T.: -0.005 min  
Response: 9795711  
Conc: 43.81 ug/ml