

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE021022AL\
 Data File : FE034973.D
 Signal(s) : FID1B.CH
 Acq On : 10 Feb 2022 19:35
 Operator : YP\AJ
 Sample : N1455-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 TP-3

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/11/2022
 Supervised By :Ankita Jodhani 02/11/2022

Integration File: sample.E
 Quant Time: Feb 11 00:42:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 020822.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 08 22:05:03 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.139	11151559	38.077 ug/mlm
Spiked Amount 50.000		Recovery =	76.15%
12) S 1-chlorooctadecane (S...	12.565	9075652	40.593 ug/mlm
Spiked Amount 50.000		Recovery =	81.19%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE021022AL\
Data File : FE034973.D
Signal(s) : FID1B.CH
Acq On : 10 Feb 2022 19:35
Operator : YP\AJ
Sample : N1455-05
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
FID_E
ClientSampleId :
TP-3


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/11/2022
Supervised By :Ankita Jodhani 02/11/2022

Integration File: sample.E
Quant Time: Feb 11 00:42:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 020822.M
Quant Title : GC Extractables
QLast Update : Tue Feb 08 22:05:03 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.139 min
Delta R.T.: 0.007 min
Response: 11151559
Conc: 38.08 ug/ml

Instrument :
FID_E
ClientSampleId :
TP-3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/11/2022
Supervised By :Ankita Jodhani 02/11/2022



#12 1-chlorooctadecane (SURR)

R.T.: 12.565 min
Delta R.T.: 0.002 min
Response: 9075652
Conc: 40.59 ug/ml m