

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE021424\
 Data File : FE046791.D
 Signal(s) : FID1B.ch
 Acq On : 15 Feb 2024 00:49
 Operator : YP\AJ
 Sample : P1187-11
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 15 05:57:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE021424.M
 Quant Title :
 QLast Update : Wed Feb 14 16:03:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.894	2648525	21.528 ug/ml
Target Compounds				
1)	N-OCTANE	1.935	664735	4.587 ug/ml
2)	N-DECANE	4.440	661292	5.081 ug/ml
3)	N-DODECANE	6.613	638297	4.842 ug/ml
4)	N-TETRADECANE	8.440	678088	5.272 ug/ml
5)	N-HEXADECANE	10.049	728399	5.523 ug/ml
6)	N-OCTADECANE	11.490	801111	5.661 ug/ml
7)	N-EICOSANE	12.799	827312	5.965 ug/ml
8)	N-DOCOSANE	13.995	817823	6.018 ug/ml
10)	N-TETRACOSANE	15.099	798393	5.936 ug/ml
11)	N-HEXACOSANE	16.120	788517	5.981 ug/ml
12)	N-OCTACOSANE	17.070	784520	6.010 ug/ml
13)	N-TRIACONTANE	17.958	796626	6.139 ug/ml
14)	N-DOTRIACONTANE	18.791	807527	6.449 ug/ml
15)	N-TETRATRIACONTANE	19.575	734489	7.136 ug/ml
16)	N-HEXATRIACONTANE	20.318	633401	7.761 ug/ml
17)	N-OCTATRIACONTANE	21.041	592057	7.709 ug/ml
18)	N-TETRACONTANE	21.920	571560	7.396 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE021424\
Data File : FE046791.D
Signal(s) : FID1B.ch
Acq On : 15 Feb 2024 00:49
Operator : YP\AJ
Sample : P1187-11
Misc :
ALS Vial : 26 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 15 05:57:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE021424.M
Quant Title :
QLast Update : Wed Feb 14 16:03:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

