

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE022223\
 Data File : FE040492.D
 Signal(s) : FID1B.ch
 Acq On : 22 Feb 2023 15:33
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 23 02:24:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE022123.M
 Quant Title :
 QLast Update : Wed Feb 22 06:24:20 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.977	6172160	46.743 ug/ml
Target Compounds			
2) N-DECANE	4.525	6632451	48.545 ug/ml
3) N-DODECANE	6.693	6787577	48.711 ug/ml
4) N-TETRADECANE	8.520	6853117	47.888 ug/ml
5) N-HEXADECANE	10.127	7049204	47.494 ug/ml
6) N-OCTADECANE	11.569	7128788	47.012 ug/ml
7) N-EICOSANE	12.878	7136932	46.927 ug/ml
8) N-DOCOSANE	14.077	6955895	46.876 ug/ml
10) N-TETRACOSANE	15.181	6919324	46.785 ug/ml
11) N-HEXACOSANE	16.203	6744348	46.594 ug/ml
12) N-OCTACOSANE	17.153	6554616	46.308 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE022223\
Data File : FE040492.D
Signal(s) : FID1B.ch
Acq On : 22 Feb 2023 15:33
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 23 02:24:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE022123.M
Quant Title :
QLast Update : Wed Feb 22 06:24:20 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

