

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE022323\
 Data File : FE040506.D
 Signal(s) : FID1B.ch
 Acq On : 23 Feb 2023 09:24
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 24 02:17:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE022123.M
 Quant Title :
 QLast Update : Wed Feb 22 06:24:20 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.977	6479148	49.068 ug/ml
Target Compounds			
1) N-OCTANE	2.010	6624066	47.719 ug/ml
2) N-DECANE	4.524	6787217	49.678 ug/ml
3) N-DODECANE	6.692	7020173	50.380 ug/ml
4) N-TETRADECANE	8.519	7154514	49.994 ug/ml
5) N-HEXADECANE	10.127	7412937	49.945 ug/ml
6) N-OCTADECANE	11.569	7534316	49.686 ug/ml
7) N-EICOSANE	12.880	7558462	49.698 ug/ml
8) N-DOCOSANE	14.078	7305827	49.235 ug/ml
10) N-TETRACOSANE	15.182	7228371	48.874 ug/ml
11) N-HEXACOSANE	16.204	7015313	48.466 ug/ml
12) N-OCTACOSANE	17.154	6798080	48.028 ug/ml
13) N-TRIACONTANE	18.042	6515081	47.511 ug/ml
14) N-DOTRIACONTANE	18.875	6073211	46.774 ug/ml
15) N-TETRATRIACONTANE	19.661	5538339	46.151 ug/ml
16) N-HEXATRIACONTANE	20.400	5067397	42.972 ug/ml
17) N-OCTATRIACONTANE	21.131	4783277	43.492 ug/ml
18) N-TETRACONTANE	22.025	4592480	41.061 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE022323\
Data File : FE040506.D
Signal(s) : FID1B.ch
Acq On : 23 Feb 2023 09:24
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 24 02:17:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE022123.M
Quant Title :
QLast Update : Wed Feb 22 06:24:20 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

