

nteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE03012
Data File : FE040612.D
Signal(s) : FID1B.ch
Acq On : 01 Mar 2023 01:43
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/01/2023
Supervised By :Ankita Jodhani 03/01/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 022823.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.281	3.251	3.364	BB	309521	3093025	78.43%	4.539%
2	4.528	4.493	4.621	BB	307290	3053067	77.42%	4.481%
3	6.230	6.191	6.377	BB	321296	3401787	86.26%	4.992%
4	6.692	6.657	6.775	BB	310627	3113462	78.95%	4.569%
5	7.325	7.293	7.496	BB	279627	3192295	80.95%	4.685%
6	8.519	8.486	8.619	BB	297461	3149353	79.86%	4.622%
7	10.126	10.079	10.228	BB	298895	3249430	82.40%	4.769%
8	11.569	11.533	11.659	BB	283866	3308280	83.89%	4.855%
9	11.885	11.853	11.967	BB	323996	3554908	90.14%	5.217%
10	12.879	12.800	12.949	BB	274005	3210921	81.42%	4.712%
11	13.323	13.279	13.412	BB	233006	2755431	69.87%	4.044%
12	13.491	13.440	13.573	BB	265280	3204355	81.25%	4.703%
13	14.077	14.011	14.157	BB	261821	3151350	79.91%	4.625%
14	15.181	15.106	15.241	BB	239408	3063113	77.67%	4.495%
15	16.203	15.778	16.268	BB	228313	2898962	73.51%	4.255%
16	17.155	17.060	17.194	VV	215776	3064996	77.72%	4.498%
17	18.043	17.742	18.083	BV	203326	2612105	66.24%	3.834%
18	18.876	18.806	18.915	BV	199901	3069052	77.82%	4.504%
19	19.661	19.607	19.695	BV	192683	2690694	68.23%	3.949%
20	20.403	20.350	20.437	BV	178354	2714148	68.82%	3.983%
21	21.134	21.094	21.165	PV	159848	2643574	67.03%	3.880%
22	22.028	21.977	22.119	VV	132939	3943669	100.00%	5.788%
Sum of corrected areas:							68137977	

Aliphatic EPH 022823.M Wed Mar 01 03:37:50 2023