

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030221\
 Data File : FE033324.D
 Signal(s) : FID1B.CH
 Acq On : 02 Mar 2021 15:07
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 03 02:10:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE022221.M
 Quant Title :
 QLast Update : Tue Feb 23 10:21:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.569	16361475	47.634 ug/ml
Target Compounds			
2) N-DECANE	4.319	18192934	54.734 ug/ml
3) N-DODECANE	6.326	18330695	53.707 ug/ml
4) N-TETRADECANE	8.119	18798440	51.608 ug/ml
5) N-HEXADECANE	9.719	18730064	51.025 ug/ml
6) N-OCTADECANE	11.161	18915965	50.028 ug/ml
7) N-EICOSANE	12.471	18796897	49.165 ug/ml
8) N-DOCOSANE	13.670	18677397	48.262 ug/ml
10) N-TETRACOSANE	14.777	18609084	47.767 ug/ml
11) N-HEXACOSANE	15.798	18273919	47.103 ug/ml
12) N-OCTACOSANE	16.749	18145495	46.301 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030221\
Data File : FE033324.D
Signal(s) : FID1B.CH
Acq On : 02 Mar 2021 15:07
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 03 02:10:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE022221.M
Quant Title :
QLast Update : Tue Feb 23 10:21:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

