

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030320\
 Data File : FE031395.D
 Signal(s) : FID1B.CH
 Acq On : 03 Mar 2020 18:05
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 04 00:32:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE021120.M
 Quant Title :
 QLast Update : Tue Feb 11 12:55:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.607	13385228	43.066 ug/ml
Target Compounds			
2) N-DECANE	4.340	15851801	47.831 ug/ml
3) N-DODECANE	6.345	15931950	48.277 ug/ml
4) N-TETRADECANE	8.141	15936375	47.158 ug/ml
5) N-HEXADECANE	9.745	16050591	46.717 ug/ml
6) N-OCTADECANE	11.192	15972453	46.318 ug/ml
7) N-EICOSANE	12.506	15784545	45.308 ug/ml
8) N-DOCOSANE	13.707	15625638	44.280 ug/ml
10) N-TETRACOSANE	14.815	15346225	43.027 ug/ml
11) N-HEXACOSANE	15.840	14920590	41.868 ug/ml
12) N-OCTACOSANE	16.792	14594561	40.278 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030320\
Data File : FE031395.D
Signal(s) : FID1B.CH
Acq On : 03 Mar 2020 18:05
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 04 00:32:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE021120.M
Quant Title :
QLast Update : Tue Feb 11 12:55:11 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

