

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\
 Method File : C44-FE030419.M
 Title :
 Last Update : Tue Mar 05 10:37:48 2019
 Response Via : Initial Calibration

Calibration Files

= S1 =FE029590.D S2 =FE029591.D
 S3 =FE029592.D S4 =FE029593.D S5 =FE029594.D

Compound		S1	S2	S3	S4	S5	Avg	%RSD	
1)	N-OCTANE	1.752	1.703	1.762	1.780	1.739	1.747	E5	1.65
2)	N-NONANE	1.799	1.744	1.800	1.811	1.764	1.783	E5	1.59
3)	N-DECANE	2.761	2.680	2.776	2.800	2.682	2.740	E5	2.02
4)	N-UNDECANE	1.861	1.808	1.871	1.871	1.810	1.844	E5	1.76
5)	N-DODECANE	1.861	1.808	1.867	1.878	1.829	1.848	E5	1.58
6)	N-TETRADECANE	1.887	1.836	1.895	1.912	1.870	1.880	E5	1.53
7)	N-HEXADECANE	1.890	1.840	1.898	1.912	1.867	1.881	E5	1.51
8)	N-OCTADECANE	1.879	1.833	1.880	1.879	1.830	1.860	E5	1.41
9)	N-EICOSANE	1.868	1.818	1.868	1.863	1.772	1.838	E5	2.32
10)	N-TETRACOSANE	1.817	1.767	1.815	1.792	1.687	1.776	E5	3.02
11)	N-OCTACOSANE	1.734	1.690	1.730	1.722	1.641	1.703	E5	2.27
12)	N-DOTRIACONTANE	1.588	1.578	1.643	1.607	1.589	1.601	E5	1.61
13) S	TETRACOSANE-d...	1.898	1.843	1.908	1.907	1.843	1.880	E5	1.79
14)	N-HEXATRIACON...	1.361	1.350	1.374	1.398	1.395	1.376	E5	1.52
15)	N-TETRACONTANE	0.886	1.135	1.090	1.073	1.080	1.053	E5	9.15
16)	N-TETRTETRACO...	4.987	8.594	8.105	7.254	7.266	7.241	E4	19.11

(#) = Out of Range