

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030822AL\
Data File : FE035248.D
Signal(s) : FID1B.CH
Acq On : 08 Mar 2022 11:21
Operator : YP\AJ
Sample : MECL2
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 09 00:03:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 022822.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 02:07:56 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
22) T n-Tetracontane (C40)	21.279	474350	7.755 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030822AL\
Data File : FE035248.D
Signal(s) : FID1B.CH
Acq On : 08 Mar 2022 11:21
Operator : YP\AJ
Sample : MECL2
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 09 00:03:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 022822.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 02:07:56 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 u1
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURRE)

R.T.: 0.000 min
Exp R.T. : 11.362 min
Response: 0
Conc: N.D.



#12 1-chlorooctadecane (SURRE)

R.T.: 0.000 min
Exp R.T. : 12.797 min
Response: 0
Conc: N.D.



#22 n-Tetracontane (C40)

R.T.: 21.279 min
Delta R.T.: -0.069 min
Response: 474350
Conc: 7.75 ug/ml