

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030823\
 Data File : FE040710.D
 Signal(s) : FID1B.ch
 Acq On : 08 Mar 2023 12:42
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 08 22:38:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE022123.M
 Quant Title :
 QLast Update : Wed Feb 22 06:24:20 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.975	6509098	49.295 ug/ml
Target Compounds			
2) N-DECANE	4.522	7724211	56.536 ug/ml
3) N-DODECANE	6.690	7807461	56.030 ug/ml
4) N-TETRADECANE	8.517	7740738	54.090 ug/ml
5) N-HEXADECANE	10.125	7792443	52.502 ug/ml
6) N-OCTADECANE	11.567	7747968	51.095 ug/ml
7) N-EICOSANE	12.877	7620504	50.106 ug/ml
8) N-DOCOSANE	14.076	7326836	49.376 ug/ml
10) N-TETRACOSANE	15.180	7237968	48.939 ug/ml
11) N-HEXACOSANE	16.202	7041357	48.646 ug/ml
12) N-OCTACOSANE	17.152	6849155	48.388 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030823\
Data File : FE040710.D
Signal(s) : FID1B.ch
Acq On : 08 Mar 2023 12:42
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 08 22:38:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE022123.M
Quant Title :
QLast Update : Wed Feb 22 06:24:20 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

