

rtres

Instrument :

FID_E

LabSampleId :

20 TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/10/2022

Supervised By :Ankita Jodhani 03/10/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE030922
Data File : FE035262.D
Signal(s) : FID1B.CH
Acq On : 09 Mar 2022 11:54
Sample : 20 TRPH STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE030922.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.184	2.162	2.223	BV	323045	2269168	99.33%	6.532%
2	4.236	4.207	4.294	M	240721	2246157	98.32%	6.466%
3	6.243	6.214	6.298	M	218886	2205170	96.53%	6.348%
4	8.028	7.996	8.127	BV	198225	2284539	100.00%	6.576%
5	9.621	9.585	9.693	BB	190954	2234138	97.79%	6.431%
6	11.055	11.002	11.128	BV	187320	2240000	98.05%	6.448%
7	12.357	12.320	12.426	BB	175584	2136264	93.51%	6.149%
8	13.550	13.510	13.610	BB	165854	2061028	90.22%	5.933%
9	14.446	14.403	14.502	BB	127159	1783366	78.06%	5.133%
10	14.647	14.605	14.710	BB	154663	1998523	87.48%	5.753%
11	15.662	15.620	15.733	BB	139298	1924934	84.26%	5.541%
12	16.608	16.563	16.679	BB	136434	1849914	80.98%	5.325%
13	17.490	17.443	17.558	BB	123928	1778553	77.85%	5.120%
14	18.319	18.271	18.388	BB	113560	1717368	75.17%	4.943%
15	19.099	19.053	19.173	BB	110171	1687820	73.88%	4.858%
16	19.834	19.788	19.911	BB	95746	1563575	68.44%	4.501%
17	20.537	20.487	20.612	BB	80701	1404850	61.49%	4.044%
18	21.350	21.288	21.430	BB	59435	1354623	59.30%	3.899%
Sum of corrected areas:						34739989		

FE030922.M Wed Mar 09 13:38:34 2022