

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE031022\
 Data File : FE035283.D
 Signal(s) : FID1B.CH
 Acq On : 10 Mar 2022 9:57
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/11/2022
 Supervised By :Ankita Jodhani 03/11/2022

Integration File: autoint1.e
 Quant Time: Mar 10 16:05:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE030922.M
 Quant Title :
 QLast Update : Wed Mar 09 13:31:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.474	4222105	47.152 ug/ml
Target Compounds			
1) N-OCTANE	2.188	5500465	47.765 ug/ml
2) N-DECANE	4.249	5554370	49.183 ug/ml
3) N-DODECANE	6.261	5396176	48.758 ug/ml
4) N-TETRADECANE	8.050	5522976	48.254 ug/ml
5) N-HEXADECANE	9.644	5419796	47.893 ug/ml
6) N-OCTADECANE	11.081	5385655	47.475 ug/ml
7) N-EICOSANE	12.386	5099087	47.231 ug/ml
8) N-DOCOSANE	13.579	4905564	46.923 ug/ml
10) N-TETRACOSANE	14.678	4741974	46.869 ug/ml
11) N-HEXACOSANE	15.694	4607069	47.252 ug/ml
12) N-OCTACOSANE	16.641	4504063	47.999 ug/ml
13) N-TRIACONTANE	17.525	4457901	49.216 ug/ml
14) N-DOTRIACONTANE	18.354	4307310	49.006 ug/mlm
15) N-TETRATRIACONTANE	19.136	4102669	47.876 ug/mlm
16) N-HEXATRIACONTANE	19.872	3804539	48.143 ug/mlm
17) N-OCTATRIACONTANE	20.575	3113332	43.797 ug/mlm
18) N-TETRACONTANE	21.404	2636839	38.345 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE031022\
Data File : FE035283.D
Signal(s) : FID1B.CH
Acq On : 10 Mar 2022 9:57
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/11/2022
Supervised By :Ankita Jodhani 03/11/2022

Integration File: autoint1.e
Quant Time: Mar 10 16:05:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE030922.M
Quant Title :
QLast Update : Wed Mar 09 13:31:22 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

