

rteres

Instrument :

FID_E

ClientSampleId :

PB143313BS

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE03142

Data File : FE035309.D

Signal(s) : FID1B.CH

Acq On : 14 Mar 2022 17:23

Sample : PB143313BS

Misc :

ALS Vial : 4 Sample Multiplier: 1

Reviewed By :Yogesh Patel 03/15/2022

Supervised By :Ankita Jodhani 03/15/2022

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 022822.M

Title : GC Extractables

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.212	3.176	3.294	BB	744585	6269196	90.00%	3.056%
2	4.255	4.214	4.274	BV	748593	6461039	92.75%	3.149%
3	4.294	4.274	4.380	VB	761326	6966037	100.00%	3.395%
4	5.817	5.778	5.925	BB	715742	6940802	99.64%	3.383%
5	6.252	6.212	6.326	BB	663119	6435673	92.39%	3.137%
6	6.868	6.820	6.982	BV	646270	6632910	95.22%	3.233%
7	8.041	7.994	8.107	BB	603280	6362121	91.33%	3.101%
8	8.141	8.110	8.237	BB	610002	6274423	90.07%	3.058%
9	8.437	8.390	8.524	BB	639392	6659320	95.60%	3.246%
10	9.218	9.172	9.323	BB	595007	6315015	90.65%	3.078%
11	9.634	9.585	9.698	BB	561217	6275331	90.08%	3.059%
12	10.612	10.545	10.656	BV	512693	5938512	85.25%	2.894%
13	10.688	10.656	10.806	VB	501279	5788323	83.09%	2.821%
14	11.070	11.023	11.128	BB	536541	6272505	90.04%	3.057%
15	11.379	11.333	11.443	BB	445918	4887248	70.16%	2.382%
16	12.374	12.285	12.395	BV	500851	5953456	85.46%	2.902%
17	12.423	12.395	12.517	VB	478084	5585801	80.19%	2.723%
18	12.718	12.672	12.757	BV	466617	5493297	78.86%	2.677%
19	12.813	12.782	12.887	VB	314470	3688827	52.95%	1.798%
20	12.982	12.934	13.052	BB	474521	5815840	83.49%	2.835%
21	13.566	13.516	13.636	BB	461464	5683469	81.59%	2.770%
22	14.594	14.538	14.613	BV	398497	5229474	75.07%	2.549%
23	14.638	14.613	14.649	VV	450614	5412326	77.70%	2.638%
24	14.664	14.649	14.763	VB	443009	5486088	78.75%	2.674%
25	15.682	15.627	15.758	BB	401614	5195548	74.58%	2.532%
26	16.146	16.086	16.162	BV	330936	5193437	74.55%	2.531%
27	16.182	16.162	16.301	VB	420337	5679169	81.53%	2.768%
28	16.526	16.464	16.590	BV	329854	4844730	69.55%	2.361%
29	16.625	16.590	16.738	VB	368799	5132212	73.67%	2.501%
30	17.513	17.453	17.607	BB	335954	4813770	69.10%	2.346%
31	17.895	17.831	17.911	BV	282831	4366196	62.68%	2.128%
32	17.936	17.911	18.098	VB	338182	5703785	81.88%	2.780%
33	18.153	18.099	18.293	BV	298452	5145745	73.87%	2.508%
34	18.339	18.293	18.458	VB	307649	4570422	65.61%	2.228%
35	19.119	19.058	19.216	BB	257157	3931405	56.44%	1.916%
36	19.854	19.805	19.920	M	200715	3139345	45.07%	1.530%

rterres
 37 20.557 20.478 20.636 M 149357 2457593 35.28% 1.198%
 38 21.375 21.323 21.500 VBA 92706 2171266 35.28% 1.198%
 Sum of corrected areas: 2051

Aliphatic EPH 022822.M Tue Mar 15 11:05:15 2022

Instrument :
 FID_E
 ClientSampleId :
 PB143313BS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/15/2022
 Supervised By :Ankita Jodhani 03/15/2022