

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE031820\
 Data File : FE031511.D
 Signal(s) : FID1B.CH
 Acq On : 18 Mar 2020 16:29
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 19 00:54:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE031020.M
 Quant Title :
 QLast Update : Wed Mar 11 09:35:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.400	18144275	49.161 ug/ml
Target Compounds			
2) N-DECANE	4.169	20734494	52.037 ug/ml
3) N-DODECANE	6.172	21201674	51.748 ug/ml
4) N-TETRADECANE	7.962	21269688	51.143 ug/ml
5) N-HEXADECANE	9.560	21519427	50.607 ug/ml
6) N-OCTADECANE	10.998	21463380	50.118 ug/ml
7) N-EICOSANE	12.307	21254642	49.721 ug/ml
8) N-DOCOSANE	13.504	21052028	49.342 ug/ml
10) N-TETRACOSANE	14.606	20811550	48.870 ug/ml
11) N-HEXACOSANE	15.625	20387595	48.540 ug/ml
12) N-OCTACOSANE	16.573	20301329	48.372 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE031820\
Data File : FE031511.D
Signal(s) : FID1B.CH
Acq On : 18 Mar 2020 16:29
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 19 00:54:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE031020.M
Quant Title :
QLast Update : Wed Mar 11 09:35:11 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

