

rteres

Instrument :
FID_E
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/21/2022
Supervised By :mohammad ahmed 03/21/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE03182
Data File : FE035464.D
Signal(s) : FID1B.CH
Acq On : 20 Mar 2022 4:29
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 98 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE030922.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.172	2.138	2.264	BB	682255	5595762	74.97%	5.328%
2	4.237	4.204	4.329	BB	645065	6280142	84.14%	5.979%
3	6.249	6.216	6.310	BV	670788	6731047	90.18%	6.409%
4	8.037	7.990	8.122	BB	670237	7129785	95.53%	6.788%
5	9.632	9.588	9.707	BB	671590	7301157	97.82%	6.952%
6	11.069	11.023	11.124	BB	651261	7463619	100.00%	7.106%
7	12.373	12.329	12.418	BB	578444	7083365	94.91%	6.744%
8	13.566	13.528	13.614	VB	551525	6648317	89.08%	6.330%
9	14.461	14.422	14.528	BB	420636	5555158	74.43%	5.289%
10	14.664	14.601	14.713	BB	485642	6131785	82.16%	5.838%
11	15.679	15.628	15.729	BB	432997	5576802	74.72%	5.310%
12	16.625	16.564	16.677	BB	394465	5200274	69.67%	4.951%
13	17.510	17.457	17.552	BB	380354	4994433	66.92%	4.755%
14	18.338	18.285	18.379	BB	354980	4943466	66.23%	4.707%
15	19.121	19.054	19.177	BB	358828	4990436	66.86%	4.751%
16	19.858	19.809	19.906	M	330639	4767019	63.87%	4.539%
17	20.560	20.497	20.662	M	281659	4426797	59.31%	4.215%
18	21.381	21.313	21.463	BB	198356	4210364	56.41%	4.009%
Sum of corrected areas:						105029728		

FE030922.M Mon Mar 21 14:54:12 2022