

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\
 Method File : C44-FE031919.M
 Title :
 Last Update : Wed Mar 20 08:08:03 2019
 Response Via : Initial Calibration

Calibration Files

= S1 =FE029729.D S2 =FE029730.D
 S3 =FE029731.D S4 =FE029732.D S5 =FE029733.D

Compound		S1	S2	S3	S4	S5	Avg	%RSD	

1)	N-OCTANE	1.328	1.269	1.248	1.447	1.273	1.313	E5	6.13
2)	N-NONANE	1.428	1.357	1.326	1.532	1.363	1.401	E5	5.86
3)	N-DECANE	2.284	2.163	2.103	2.421	2.177	2.229	E5	5.61
4)	N-UNDECANE	1.597	1.512	1.469	1.687	1.531	1.559	E5	5.46
5)	N-DODECANE	1.649	1.566	1.524	1.748	1.600	1.617	E5	5.32
6)	N-TETRADECANE	1.747	1.676	1.651	1.890	1.775	1.748	E5	5.39
7)	N-HEXADECANE	1.787	1.729	1.725	1.977	1.934	1.831	E5	6.45
8)	N-OCTADECANE	1.780	1.729	1.743	2.001	2.088	1.868	E5	8.83
9)	N-EICOSANE	1.727	1.674	1.678	1.899	2.100	1.816	E5	10.12
10)	N-TETRACOSANE	1.480	1.412	1.399	1.574	1.965	1.566	E5	14.91
11)	N-OCTACOSANE	1.374	1.320	1.319	1.547	1.666	1.445	E5	10.71
12)	N-DOTRIACONTANE	1.449	1.421	1.491	1.874	1.957	1.639	E5	15.61
13) S	TETRACOSANE-d...	1.565	1.495	1.486	1.666	2.081	1.659	E5	14.89
14)	N-HEXATRIACON...	1.402	1.362	1.334	1.564	1.566	1.445	E5	7.73
15)	N-TETRACONTANE	1.176	1.260	1.127	1.336	1.326	1.245	E5	7.36
16)	N-TETRTETRACO...	0.866	1.062	0.884	1.053	1.023	0.978	E5	9.70

(#) = Out of Range