

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE031920\
 Data File : FE031525.D
 Signal(s) : FID1B.CH
 Acq On : 19 Mar 2020 18:11
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 20 01:07:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE031020.M
 Quant Title :
 QLast Update : Wed Mar 11 09:35:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.428	17678034	47.898 ug/ml
Target Compounds			
2) N-DECANE	4.182	20872782	52.384 ug/ml
3) N-DODECANE	6.188	21423673	52.290 ug/ml
4) N-TETRADECANE	7.981	21403347	51.465 ug/ml
5) N-HEXADECANE	9.581	21572894	50.733 ug/ml
6) N-OCTADECANE	11.022	21353467	49.861 ug/ml
7) N-EICOSANE	12.330	21007064	49.142 ug/ml
8) N-DOCOSANE	13.530	20674126	48.457 ug/ml
10) N-TETRACOSANE	14.634	20308742	47.689 ug/ml
11) N-HEXACOSANE	15.655	19846491	47.251 ug/ml
12) N-OCTACOSANE	16.605	19760947	47.084 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE031920\
Data File : FE031525.D
Signal(s) : FID1B.CH
Acq On : 19 Mar 2020 18:11
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 20 01:07:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE031020.M
Quant Title :
QLast Update : Wed Mar 11 09:35:11 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

