

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032323AL\
 Data File : FE040989.D
 Signal(s) : FID1B.ch
 Acq On : 24 Mar 2023 00:16
 Operator : YP\AJ
 Sample : 02028-01DL 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 TR-05-032123DL

Integration File: autoint1.e
 Quant Time: Mar 24 02:56:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 022823.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 28 03:05:09 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.869	1366725	7.316 ug/ml
Spiked Amount	50.000	Recovery	= 14.63%
12) S 1-chlorooctadecane (S...	13.303	17164759	115.975 ug/ml
Spiked Amount	50.000	Recovery	= 231.95%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032323AL\
Data File : FE040989.D
Signal(s) : FID1B.ch
Acq On : 24 Mar 2023 00:16
Operator : YP\AJ
Sample : 02028-01DL 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
TR-05-032123DL

Integration File: autoint1.e
Quant Time: Mar 24 02:56:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 022823.M
Quant Title : GC Extractables
QLast Update : Tue Feb 28 03:05:09 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

