

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032421\
 Data File : FE033437.D
 Signal(s) : FID1B.CH
 Acq On : 24 Mar 2021 10:11
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 25 01:09:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE031621.M
 Quant Title :
 QLast Update : Wed Mar 17 09:27:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.746	19344779	50.336 ug/ml
Target Compounds			
1) N-OCTANE	2.432	20298532	51.309 ug/ml
2) N-DECANE	4.490	21154921	51.427 ug/ml
3) N-DODECANE	6.504	21451887	51.456 ug/ml
4) N-TETRADECANE	8.299	22009782	51.412 ug/ml
5) N-HEXADECANE	9.899	22024762	51.285 ug/ml
6) N-OCTADECANE	11.342	22393952	51.207 ug/ml
7) N-EICOSANE	12.651	22298871	50.867 ug/ml
8) N-DOCOSANE	13.848	22184877	50.544 ug/ml
10) N-TETRACOSANE	14.953	22076937	50.049 ug/ml
11) N-HEXACOSANE	15.972	21649458	49.609 ug/ml
12) N-OCTACOSANE	16.920	21360604	49.021 ug/ml
13) N-TRIACONTANE	17.808	20807435	48.314 ug/ml
14) N-DOTRIACONTANE	18.639	19809948	46.868 ug/ml
15) N-TETRATRIACONTANE	19.424	18459193	48.127 ug/ml
16) N-HEXATRIACONTANE	20.162	16677987	48.149 ug/ml
17) N-OCTATRIACONTANE	20.908	14892710	47.064 ug/ml
18) N-TETRACONTANE	21.840	14069900	48.206 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032421\
Data File : FE033437.D
Signal(s) : FID1B.CH
Acq On : 24 Mar 2021 10:11
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 25 01:09:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE031621.M
Quant Title :
QLast Update : Wed Mar 17 09:27:02 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

