

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032423AL\
 Data File : FE041000.D
 Signal(s) : FID1B.ch
 Acq On : 24 Mar 2023 13:09
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Mar 25 00:58:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 022823.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 28 03:05:09 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
19) T n-Tetratriacontane (C34)	19.674	23500	0.152 ug/ml
20) T n-Hexatriacontane (C36)	20.439	49668	0.342 ug/ml
22) T n-Tetracontane (C40)	22.019	294215	1.904 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032423AL\
Data File : FE041000.D
Signal(s) : FID1B.ch
Acq On : 24 Mar 2023 13:09
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 25 00:58:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 022823.M
Quant Title : GC Extractables
QLast Update : Tue Feb 28 03:05:09 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 0.000 min
Exp R.T.: 11.879 min
Response: 0
Conc: N.D.

Instrument :
FID_E
ClientSampleId :




#12 1-chlorooctadecane (SURR)

R.T.: 0.000 min
Exp R.T.: 13.318 min
Response: 0
Conc: N.D.




#19 n-Tetratriacontane (C34)

R.T.: 19.674 min
Delta R.T.: 0.023 min
Response: 23500
Conc: 0.15 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 20.439 min
Delta R.T.: 0.046 min
Response: 49668
Conc: 0.34 ug/ml



#22 n-Tetracontane (C40)

R.T.: 22.019 min
Delta R.T.: 0.000 min
Response: 294215
Conc: 1.90 ug/ml

Instrument :

FID_E

ClientSampleId :