

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032519\
 Data File : FE029755.D
 Signal(s) : FID1B.CH
 Acq On : 25 Mar 2019 14:24
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 26 00:52:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE031919.M
 Quant Title :
 QLast Update : Wed Mar 20 08:08:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.655	8913737	53.744 ug/ml
Target Compounds			
1) N-OCTANE	2.399	67501342	514.144 ug/ml
2) N-NONANE	3.378	68979925	492.329 ug/ml
3) N-DECANE	4.438	105542521	473.401 ug/ml
4) N-UNDECANE	5.468	106117080	680.614 ug/ml
5) N-DODECANE	6.450	106938117	661.134 ug/ml
6) N-TETRADECANE	8.250	107774091	616.561 ug/ml
7) N-HEXADECANE	9.849	89850837	490.829 ug/ml
8) N-OCTADECANE	11.273	44963702	240.684 ug/ml
9) N-EICOSANE	12.564	17877901	98.468 ug/ml
10) N-TETRACOSANE	14.871	17661739	112.783 ug/ml
11) N-OCTACOSANE	16.834	8681478	60.068 ug/ml
12) N-DOTRIACONTANE	18.555	7708509	47.043 ug/ml
14) N-HEXATRIACONTANE	20.075	4797955	33.193 ug/mlm
15) N-TETRACONTANE	21.693	3815603	30.647 ug/mlm
16) N-TETRTETRACOOCTANE	24.377	3120451	31.913 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032519\
Data File : FE029755.D
Signal(s) : FID1B.CH
Acq On : 25 Mar 2019 14:24
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 26 00:52:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE031919.M
Quant Title :
QLast Update : Wed Mar 20 08:08:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

