

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE032525AL\  
 Data File : FE052966.D  
 Signal(s) : FID1B.ch  
 Acq On : 25 Mar 2025 19:56  
 Operator : YP\AJ  
 Sample : Q1635-02  
 Misc :  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :  
 TP-2-EPH

Integration File: autoint1.e  
 Quant Time: Mar 26 03:27:41 2025  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 030325.M  
 Quant Title : GC Extractables  
 QLast Update : Mon Mar 03 13:39:39 2025  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response | Conc Units   |
|--------------------------------|--------|----------|--------------|
| -----                          |        |          |              |
| System Monitoring Compounds    |        |          |              |
| 9) S ortho-Terphenyl (SURR)    | 11.907 | 4500018  | 26.693 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 53.39%     |
| 12) S 1-chlorooctadecane (S... | 13.352 | 3390269  | 27.920 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 55.84%     |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE032525AL\  
Data File : FE052966.D  
Signal(s) : FID1B.ch  
Acq On : 25 Mar 2025 19:56  
Operator : YP\AJ  
Sample : Q1635-02  
Misc :  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :  
TP-2-EPH

Integration File: autoint1.e  
Quant Time: Mar 26 03:27:41 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 030325.M  
Quant Title : GC Extractables  
QLast Update : Mon Mar 03 13:39:39 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.907 min  
Delta R.T.: -0.001 min  
Response: 4500018  
Conc: 26.69 ug/ml

Instrument :  
FID\_E  
ClientSampleId :  
TP-2-EPH



#12 1-chlorooctadecane (SURR)

R.T.: 13.352 min  
Delta R.T.: -0.002 min  
Response: 3390269  
Conc: 27.92 ug/ml