

rteres

Area Percent Report

Instrument :
FID_E
LabSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

mohammad
4/1/2020 8:25:24 AM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE032720\
Data File : FE031609.D
Signal(s) : FID1B.CH
Acq On : 27 Mar 2020 14:12
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE031020.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.144	2.085	2.283	BV	2206595	18038002	87.82%	5.572%
2	4.161	4.082	4.292	BB	1839473	19139038	93.18%	5.912%
3	6.165	6.062	6.318	BB	1692853	19918024	96.97%	6.152%
4	7.956	7.865	8.085	BB	1610227	20163323	98.16%	6.228%
5	9.553	9.442	9.678	BB	1703571	20509251	99.85%	6.335%
6	10.992	10.817	11.105	PB	1580231	20540261	100.00%	6.344%
7	12.301	12.198	12.415	BB	1486031	20433852	99.48%	6.312%
8	13.497	13.375	13.612	BB	1464492	20326888	98.96%	6.279%
9	14.395	14.292	14.502	BB	1148273	17489766	85.15%	5.402%
10	14.600	14.515	14.735	BB	1362474	20213419	98.41%	6.244%
11	15.620	15.505	15.752	BB	1251074	19782207	96.31%	6.110%
12	16.568	16.435	16.682	BB	1357965	19619022	95.51%	6.060%
13	17.455	17.232	17.548	BB	1209983	19479822	94.84%	6.017%
14	18.287	18.062	18.358	BB	1134787	18455741	89.85%	5.701%
15	19.068	18.778	19.122	BB	1007257	16288376	79.30%	5.031%
16	19.806	19.749	19.843	M	823365	13230823	64.41%	4.087%
17	20.509	20.451	20.540	M	616047	10621638	51.71%	3.281%
18	21.313	21.195	21.398	BB	416538	9499877	46.25%	2.934%
Sum of corrected areas:						323749329		

FE031020.M Tue Mar 31 12:50:31 2020