

rteres

Area Percent Report

Instrument :
FID_E
LabSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
4/6/2021 2:50:03 PM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE040521\
Data File : FE033481.D
Signal(s) : FID1B.CH
Acq On : 05 Apr 2021 21:44
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040521.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.428	2.391	2.507	BB	2659823	20630100	94.08%	5.623%
2	4.484	4.442	4.526	BB	2150640	21363101	97.42%	5.822%
3	6.498	6.450	6.538	BB	1873846	21500161	98.05%	5.860%
4	8.295	8.240	8.318	BB	1860337	21901402	99.88%	5.969%
5	9.896	9.844	9.925	BB	1783064	21759060	99.23%	5.930%
6	11.338	11.282	11.372	BB	1757089	21928225	100.00%	5.976%
7	12.648	12.590	12.693	BB	1659435	21697527	98.95%	5.914%
8	13.847	13.787	13.882	BB	1639320	21491899	98.01%	5.858%
9	14.743	14.685	14.808	BB	1281971	18787359	85.68%	5.120%
10	14.949	14.893	14.997	BB	1522312	21409371	97.63%	5.835%
11	15.970	15.908	16.018	BB	1466416	21104019	96.24%	5.752%
12	16.920	16.855	16.971	BB	1419176	21032057	95.91%	5.732%
13	17.808	17.744	17.858	BB	1393899	20786005	94.79%	5.665%
14	18.640	18.578	18.700	BB	1362033	20426622	93.15%	5.567%
15	19.425	19.355	19.481	BB	1256583	19573263	89.26%	5.335%
16	20.163	20.095	20.188	BV	1159000	18268473	83.31%	4.979%
17	20.909	20.834	20.972	BB	880177	16703896	76.18%	4.553%
18	21.844	21.749	21.883	M	630698	16545577	75.45%	4.509%
Sum of corrected areas:							366908117	

FE040521.M Tue Apr 06 09:38:56 2021