

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE040720\
 Data File : FE031704.D
 Signal(s) : FID1B.CH
 Acq On : 07 Apr 2020 14:37
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Sohil
 4/9/2020 10:15:33 AM

Integration File: autoint1.e
 Quant Time: Apr 07 15:05:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
 Quant Title :
 QLast Update : Tue Apr 07 15:01:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.367	1859091	5.131 ug/ml
Target Compounds			
1) N-OCTANE	2.159	1947244	4.964 ug/ml
2) N-DECANE	4.147	1873949	4.813 ug/ml
3) N-DODECANE	6.146	1542119	3.942 ug/mlm
4) N-TETRADECANE	7.934	1675633	4.070 ug/mlm
5) N-HEXADECANE	9.530	2018346	4.937 ug/ml
6) N-OCTADECANE	10.968	2122366	4.998 ug/ml
7) N-EICOSANE	12.273	2123629	5.057 ug/ml
8) N-DOCOSANE	13.470	2158293	5.113 ug/ml
10) N-TETRACOSANE	14.570	2130557	5.113 ug/ml
11) N-HEXACOSANE	15.589	2109818	5.114 ug/ml
12) N-OCTACOSANE	16.539	2112984	5.171 ug/ml
13) N-TRIACONTANE	17.424	2135248	5.242 ug/ml
14) N-DOTRIACONTANE	18.255	2052228	5.277 ug/ml
15) N-TETRATRIACONTANE	19.038	1920468	5.236 ug/ml
16) N-HEXATRIACONTANE	19.775	1798662	5.273 ug/ml
17) N-OCTATRIACONTANE	20.477	1569065	5.092 ug/ml
18) N-TETRACONTANE	21.276	1674707	5.379 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE040720\
Data File : FE031704.D
Signal(s) : FID1B.CH
Acq On : 07 Apr 2020 14:37
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
5 TRPH STD

Manual Integrations
APPROVED

Sohil
4/9/2020 10:15:33 AM

Integration File: autoint1.e
Quant Time: Apr 07 15:05:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
Quant Title :
QLast Update : Tue Apr 07 15:01:16 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

