

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE040723\
 Data File : FE041252.D
 Signal(s) : FID1B.ch
 Acq On : 08 Apr 2023 08:19
 Operator : YP\AJ
 Sample : 02213-08
 Misc : 5 PPM WATER LOQ
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 LOQ-WATER-02-QT2-2023

Integration File: autoint1.e
 Quant Time: Apr 10 02:16:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040523.M
 Quant Title :
 QLast Update : Thu Apr 06 02:04:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.929	2705257	21.254 ug/ml
Target Compounds			
2) N-DECANE	4.485	810534	5.465 ug/ml
3) N-DODECANE	6.650	833687	5.495 ug/ml
4) N-TETRADECANE	8.475	858413	5.625 ug/ml
5) N-HEXADECANE	10.081	872720	5.666 ug/ml
6) N-OCTADECANE	11.523	873237	5.708 ug/ml
7) N-EICOSANE	12.831	891639	5.912 ug/ml
8) N-DOCOSANE	14.029	852314	5.894 ug/ml
10) N-TETRACOSANE	15.132	846998	5.931 ug/ml
11) N-HEXACOSANE	16.153	826011	5.938 ug/ml
12) N-OCTACOSANE	17.104	820165	5.970 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE040723\
Data File : FE041252.D
Signal(s) : FID1B.ch
Acq On : 08 Apr 2023 08:19
Operator : YP\AJ
Sample : 02213-08
Misc : 5 PPM WATER LOQ
ALS Vial : 44 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
LOQ-WATER-02-QT2-2023

Integration File: autoint1.e
Quant Time: Apr 10 02:16:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040523.M
Quant Title :
QLast Update : Thu Apr 06 02:04:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

