

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/09/2025
Supervised By :mohammad ahmed 04/10/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE04082
Data File : FE053237.D
Signal(s) : FID1B.ch
Acq On : 08 Apr 2025 12:39
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 032925.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.237	3.204	3.324	BB	281897	2876121	91.24%	4.723%
2	4.499	4.468	4.581	BB	294607	2998049	95.11%	4.923%
3	6.212	6.171	6.324	BB	298650	3152240	100.00%	5.176%
4	6.679	6.648	6.754	BB	287913	2984777	94.69%	4.901%
5	7.315	7.278	7.438	BB	276477	2999383	95.15%	4.925%
6	8.516	8.481	8.604	VB	268764	2932933	93.04%	4.816%
7	10.130	10.101	10.208	BB	261930	2964626	94.05%	4.868%
8	11.580	11.541	11.654	BB	246328	2920783	92.66%	4.796%
9	11.895	11.838	11.935	BV	269881	3013612	95.60%	4.948%
10	12.894	12.854	12.961	BB	225318	2741248	86.96%	4.501%
11	13.340	13.284	13.421	VB	188741	2320432	73.61%	3.810%
12	13.508	13.471	13.574	BB	214399	2659699	84.37%	4.367%
13	14.096	14.061	14.161	VB	208363	2639331	83.73%	4.334%
14	15.202	15.121	15.271	BB	198068	2652420	84.14%	4.355%
15	16.227	16.144	16.294	BB	203920	2639124	83.72%	4.334%
16	17.179	16.998	17.251	BB	197477	2716470	86.18%	4.461%
17	18.069	18.001	18.124	BB	208761	2945453	93.44%	4.837%
18	18.906	18.851	18.964	BB	198051	2989005	94.82%	4.908%
19	19.690	19.628	19.758	BV	178964	2686858	85.24%	4.412%
20	20.431	20.371	20.484	BB	146083	2332400	73.99%	3.830%
21	21.171	21.101	21.251	BB	116617	2279520	72.31%	3.743%
22	22.088	21.664	22.238	BB	84111	2455061	77.88%	4.031%
Sum of corrected areas:						60899545		

Aliphatic EPH 032925.M Wed Apr 09 06:07:36 2025