

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\
 Method File : C44-FE040919.M
 Title :
 Last Update : Wed Apr 10 07:42:44 2019
 Response Via : Initial Calibration

Calibration Files

= S1 =FE029821.D S2 =FE029822.D
 S3 =FE029823.D S4 =FE029824.D S5 =FE029825.D

Compound		S1	S2	S3	S4	S5	Avg		%RSD

1)	N-OCTANE	1.912	1.855	1.923	1.969	2.006	1.933	E5	2.97
2)	N-NONANE	1.968	1.901	1.966	2.008	2.049	1.978	E5	2.79
3)	N-DECANE	3.017	2.917	3.031	3.100	3.155	3.044	E5	2.96
4)	N-UNDECANE	2.030	1.964	2.040	2.089	2.129	2.050	E5	3.06
5)	N-DODECANE	2.028	1.964	2.040	2.092	2.133	2.051	E5	3.14
6)	N-TETRADECANE	2.051	1.991	2.073	2.130	2.172	2.083	E5	3.38
7)	N-HEXADECANE	2.048	1.994	2.078	2.142	2.186	2.090	E5	3.63
8)	N-OCTADECANE	2.035	1.984	2.066	2.133	2.178	2.079	E5	3.72
9)	N-EICOSANE	2.026	1.976	2.060	2.130	2.171	2.072	E5	3.78
10)	N-TETRACOSANE	2.002	1.952	2.034	2.099	2.148	2.047	E5	3.79
11)	N-OCTACOSANE	1.969	1.934	2.017	2.082	2.159	2.032	E5	4.43
12)	N-DOTRIACONTANE	1.896	1.871	1.957	2.018	2.137	1.976	E5	5.39
13) S	TETRACOSANE-d...	2.081	2.039	2.112	2.181	2.237	2.130	E5	3.72
14)	N-HEXATRIACON...	1.769	1.751	1.805	1.868	1.887	1.816	E5	3.29
15)	N-TETRACONTANE	1.563	1.613	1.692	1.681	1.650	1.640	E5	3.22
16)	N-TETRTRETRACO...	1.253	1.400	1.477	1.469	1.473	1.414	E5	6.76

(#) = Out of Range