

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\
 Method File : FF041320.M
 Title :
 Last Update : Tue Apr 14 11:30:16 2020
 Response Via : Initial Calibration

Calibration Files

50 =FF006801.D 100 =FF006800.D 20 =FF006802.D
 10 =FF006803.D 5 =FF006804.D

	Compound	50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	1.275	1.373	1.317	1.289	1.144	1.280 E5	6.63
2)	N-DECANE	1.289	1.382	1.325	1.283	1.117	1.279 E5	7.71
3)	N-DODECANE	1.329	1.412	1.366	1.318	1.130	1.311 E5	8.20
4)	N-TETRADECANE	1.351	1.425	1.386	1.317	1.091	1.314 E5	9.95
5)	N-HEXADECANE	1.348	1.410	1.398	1.351	1.188	1.339 E5	6.62
6)	N-OCTADECANE	1.404	1.464	1.466	1.431	1.302	1.414 E5	4.78
7)	N-EICOSANE	1.381	1.440	1.450	1.419	1.300	1.398 E5	4.34
8)	N-DOCOSANE	1.392	1.449	1.464	1.438	1.330	1.415 E5	3.86
9) S	TETRACOSANE-d50 (S...	1.200	1.248	1.263	1.241	1.148	1.220 E5	3.79
10)	N-TETRACOSANE	1.378	1.432	1.452	1.430	1.330	1.405 E5	3.55
11)	N-HEXACOSANE	1.363	1.412	1.439	1.422	1.329	1.393 E5	3.27
12)	N-OCTACOSANE	1.347	1.388	1.420	1.420	1.351	1.385 E5	2.56
13)	N-TRIACONTANE	1.340	1.371	1.412	1.437	1.378	1.387 E5	2.72
14)	N-DOTRIACONTANE	1.298	1.309	1.361	1.425	1.446	1.368 E5	4.88
15)	N-TETRATRIACONTANE	1.186	1.180	1.253	1.271	1.179	1.214 E5	3.66
16)	N-HEXATRIACONTANE	1.187	1.127	1.267	1.365	1.140	1.217 E5	8.13
17)	N-OCTATRIACONTANE	0.968	0.890	1.071	1.047	0.939	0.983 E5	7.65
18)	N-TETRACONTANE	1.116	0.986	1.194	1.407	1.827	1.306 E5	25.18

(#) = Out of Range