

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041019\
 Data File : FE029847.D
 Signal(s) : FID1B.CH
 Acq On : 10 Apr 2019 17:36
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Apr 11 08:07:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE041019.M
 Quant Title :
 QLast Update : Wed Apr 10 08:20:21 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.643	10251252	50.532 ug/ml
Target Compounds			
2) N-DECANE	4.385	11587443	52.731 ug/ml
3) N-DODECANE	6.393	11748972	52.452 ug/ml
4) N-TETRADECANE	8.188	11937721	51.694 ug/ml
5) N-HEXADECANE	9.790	12051051	51.255 ug/ml
6) N-OCTADECANE	11.232	12116747	51.029 ug/ml
7) N-EICOSANE	12.544	11989555	50.601 ug/ml
8) N-DOCOSANE	13.744	11874644	50.565 ug/ml
10) N-TETRACOSANE	14.849	11722201	50.509 ug/ml
11) N-HEXACOSANE	15.871	11625754	50.416 ug/ml
12) N-OCTACOSANE	16.822	11582381	50.301 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041019\
Data File : FE029847.D
Signal(s) : FID1B.CH
Acq On : 10 Apr 2019 17:36
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Apr 11 08:07:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE041019.M
Quant Title :
QLast Update : Wed Apr 10 08:20:21 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

