

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\
 Method File : FE041019.M
 Title :
 Last Update : Wed Apr 10 08:20:21 2019
 Response Via : Initial Calibration

Calibration Files

50 =FE029830.D 100 =FE029829.D 20 =FE029831.D
 10 =FE029832.D 5 =FE029833.D

Compound		50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	2.146	2.204	2.197	2.351	2.266	2.233 E5	3.51
2)	N-DECANE	2.167	2.235	2.189	2.288	2.108	2.197 E5	3.11
3)	N-DODECANE	2.194	2.266	2.222	2.344	2.173	2.240 E5	3.02
4)	N-TETRADECANE	2.230	2.292	2.268	2.434	2.323	2.309 E5	3.36
5)	N-HEXADECANE	2.248	2.308	2.309	2.483	2.409	2.351 E5	3.98
6)	N-OCTADECANE	2.255	2.312	2.330	2.503	2.473	2.374 E5	4.54
7)	N-EICOSANE	2.233	2.285	2.314	2.507	2.508	2.369 E5	5.46
8)	N-DOCOSANE	2.211	2.268	2.297	2.485	2.481	2.348 E5	5.39
9) S	TETRACOSANE-d50 (S...	1.910	1.961	1.979	2.148	2.145	2.029 E5	5.45
10)	N-TETRACOSANE	2.187	2.243	2.264	2.459	2.451	2.321 E5	5.42
11)	N-HEXACOSANE	2.173	2.226	2.246	2.443	2.442	2.306 E5	5.53
12)	N-OCTACOSANE	2.164	2.212	2.237	2.435	2.464	2.303 E5	5.96
13)	N-TRIACONTANE	2.155	2.197	2.223	2.424	2.493	2.299 E5	6.53
14)	N-DOTRIACONTANE	2.089	2.122	2.150	2.330	2.427	2.224 E5	6.63
15)	N-TETRATRIACONTANE	2.042	2.087	2.110	2.282	2.300	2.164 E5	5.47
16)	N-HEXATRIACONTANE	1.979	2.013	2.043	2.209	2.144	2.078 E5	4.62
17)	N-OCTATRIACONTANE	1.789	1.847	1.846	1.990	1.970	1.888 E5	4.61
18)	N-TETRACONTANE	1.803	1.856	1.859	2.032	1.955	1.901 E5	4.82

(#) = Out of Range