

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/14/2025
Supervised By :mohammad ahmed 04/15/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE04112
Data File : FE053321.D
Signal(s) : FID1B.ch
Acq On : 11 Apr 2025 14:11
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 032925.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.235	3.198	3.311	BB	260507	2648233	77.67%	4.401%
2	4.496	4.461	4.588	BB	270955	2811224	82.45%	4.671%
3	6.210	6.178	6.331	BB	278169	2947883	86.46%	4.898%
4	6.677	6.638	6.774	BB	272002	2835177	83.15%	4.711%
5	7.312	7.268	7.441	BB	255061	2811061	82.44%	4.671%
6	8.513	8.474	8.628	BB	256932	2790935	81.85%	4.638%
7	10.127	10.091	10.214	PB	245852	2826079	82.88%	4.696%
8	11.575	11.534	11.648	BB	235054	2774374	81.37%	4.610%
9	11.890	11.834	11.934	BV	258973	2845075	83.44%	4.728%
10	12.889	12.848	12.968	BB	213333	2608548	76.50%	4.335%
11	13.335	13.288	13.428	BB	178729	2230764	65.42%	3.707%
12	13.503	13.464	13.584	BB	204618	2513355	73.71%	4.176%
13	14.091	14.004	14.168	BB	207619	2608774	76.51%	4.335%
14	15.197	15.131	15.311	BB	204609	2688977	78.86%	4.468%
15	16.221	16.118	16.324	BB	202227	2735741	80.23%	4.546%
16	17.173	17.088	17.244	BB	207769	2889265	84.74%	4.801%
17	18.063	18.001	18.108	BB	230458	3206361	94.04%	5.328%
18	18.897	18.851	18.958	BB	223918	3409716	100.00%	5.666%
19	19.681	19.618	19.751	PV	208244	3102701	91.00%	5.156%
20	20.423	20.371	20.511	BV	171451	2753695	80.76%	4.576%
21	21.161	21.101	21.228	BV	119490	2286754	67.07%	3.800%
22	22.073	21.961	22.194	BB	73475	1854740	54.40%	3.082%
Sum of corrected areas:							60179432	

Aliphatic EPH 032925.M Sat Apr 12 01:15:22 2025