

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041620\
 Data File : FE031746.D
 Signal(s) : FID1B.CH
 Acq On : 16 Apr 2020 14:45
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Apr 17 01:55:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
 Quant Title :
 QLast Update : Tue Apr 07 15:06:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.396	16675765	45.788 ug/ml
Target Compounds			
2) N-DECANE	4.160	18911491	48.939 ug/ml
3) N-DODECANE	6.163	19215345	51.294 ug/ml
4) N-TETRADECANE	7.954	20072405	50.641 ug/ml
5) N-HEXADECANE	9.552	19647943	48.180 ug/ml
6) N-OCTADECANE	10.992	20239432	47.665 ug/ml
7) N-EICOSANE	12.301	19669015	46.732 ug/ml
8) N-DOCOSANE	13.499	19547078	46.102 ug/ml
10) N-TETRACOSANE	14.603	19070901	45.560 ug/ml
11) N-HEXACOSANE	15.623	18699002	45.121 ug/ml
12) N-OCTACOSANE	16.572	18459034	44.868 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041620\
Data File : FE031746.D
Signal(s) : FID1B.CH
Acq On : 16 Apr 2020 14:45
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Apr 17 01:55:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
Quant Title :
QLast Update : Tue Apr 07 15:06:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

