

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041620\
 Data File : FE031765.D
 Signal(s) : FID1B.CH
 Acq On : 17 Apr 2020 6:36
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Apr 17 08:17:28 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
 Quant Title :
 QLast Update : Tue Apr 07 15:06:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.396	18145887	49.824 ug/ml
Target Compounds			
2) N-DECANE	4.160	20411489	52.821 ug/ml
3) N-DODECANE	6.162	20774426	55.455 ug/ml
4) N-TETRADECANE	7.954	21697831	54.742 ug/ml
5) N-HEXADECANE	9.551	21298164	52.227 ug/ml
6) N-OCTADECANE	10.992	21950094	51.693 ug/ml
7) N-EICOSANE	12.300	21369976	50.774 ug/ml
8) N-DOCOSANE	13.498	21261913	50.147 ug/ml
10) N-TETRACOSANE	14.602	20764498	49.605 ug/ml
11) N-HEXACOSANE	15.623	20330440	49.058 ug/ml
12) N-OCTACOSANE	16.573	20081061	48.811 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041620\
Data File : FE031765.D
Signal(s) : FID1B.CH
Acq On : 17 Apr 2020 6:36
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Apr 17 08:17:28 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
Quant Title :
QLast Update : Tue Apr 07 15:06:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

