

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041720\
 Data File : FE031769.D
 Signal(s) : FID1B.CH
 Acq On : 17 Apr 2020 10:24
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 18 00:50:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
 Quant Title :
 QLast Update : Tue Apr 07 15:06:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.390	17729550	48.681 ug/ml
Target Compounds			
2) N-DECANE	4.157	19662313	50.882 ug/ml
3) N-DODECANE	6.159	19970711	53.310 ug/ml
4) N-TETRADECANE	7.949	21039111	53.080 ug/ml
5) N-HEXADECANE	9.547	20757366	50.900 ug/ml
6) N-OCTADECANE	10.987	21342790	50.263 ug/ml
7) N-EICOSANE	12.294	20803849	49.429 ug/ml
8) N-DOCOSANE	13.492	20736890	48.908 ug/ml
10) N-TETRACOSANE	14.594	20308747	48.517 ug/ml
11) N-HEXACOSANE	15.615	19961392	48.167 ug/ml
12) N-OCTACOSANE	16.565	19745374	47.995 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE041720\
Data File : FE031769.D
Signal(s) : FID1B.CH
Acq On : 17 Apr 2020 10:24
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 18 00:50:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE040720.M
Quant Title :
QLast Update : Tue Apr 07 15:06:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

